









Use this chart to determine whether Omnilux Clear is suitable for you

ACNE TYPE	DESCRIPTION	SUITABLE FOR OMNILUX CLEAR™
CLEAR SKIN	 <p>No pimples or spots or yellow pus-filled spots. No whiteheads or blackheads.</p>	 Omnilux Clear is NOT indicated for clear skin. Omnilux Clear should NOT be used to prevent pimples or spots forming.
MILD ACNE	 <p>A few pimples or spots, a few pus-filled spots, some whiteheads or blackheads</p>	 Mild acne is suitable for treatment with Omnilux Clear.
MODERATE ACNE	 <p>Some pimples or spots, some pus-filled spots, only a few small nodules. Many whiteheads or blackheads.</p>	 Moderate acne is suitable for treatment with Omnilux Clear.
SEVERE ACNE	 <p>Many pimples or spots, many pus-filled spots. Some nodules. Many whiteheads or blackheads.</p>	 Omnilux Clear is NOT indicated for severe acne. Please consult your doctor for a more appropriate treatment.
WHITEHEADS OR BLACKHEADS (COMEDONAL ACNE)	 <p>Many whiteheads or blackheads. No pimples or spots. No pus-filled spots.</p>	 Omnilux Clear is NOT indicated for whiteheads or blackheads. Please consult your doctor for a more appropriate treatment.

Contraindications

DO NOT use Omnilux Clear for any other conditions other than mild to moderate acne vulgaris (detailed in the above chart). The Omnilux Clear device has not been tested for any other conditions than those listed, and the risk is unknown. DO NOT use the Omnilux Clear device if you suffer from any photosensitive disorder (sensitization to light) of the skin or eyes OR if you are taking any medication that can cause photosensitivity. If you use the Omnilux Clear device and you are sensitive to light or are taking any medication that can cause photosensitivity you may cause a severe reaction.

SEE FULL LIST

Warnings

Some people are sensitive to light. Before using Omnilux Clear to treat your acne vulgaris you must carry out a sensitivity test on your inner arm and wait 6 hours. If after purchasing Omnilux Clear you find that you are photosensitive please return the device to the retailer within 45 days of purchase.

What to expect

People who have undergone a course of 4 weekly treatments for 6 weeks using Omnilux Clear report a reduction of mild to moderate acne vulgaris.

Potential risks

Immediately after treatment skin may appear red. Skin should return to normal within 1-2 hours. Some users have reported pigmentation, skin flaking or peeling, associated itchiness and dryness. This is short lived and clears within a few days.

Temporary headaches immediately following the light treatment. Headaches can be treated with an appropriate over the counter analgesic and rehydration.

This list although not exhaustive lists the most reported side effects.